



Product Service

Attestation of Conformity

No. N8A 083766 0007 Rev. 01

Holder of Attestation: Jinan Biobase Biotech Co., Ltd.

North Side of Jiwang Road
Mingshui Economic Development District
250200 Zhangqiu, Jinan City
PEOPLE'S REPUBLIC OF CHINA

**Product: Laboratory Equipment
(Microbiological Safety Cabinet)****Brand: BIOBASE****Model(s): BSC-3FA2, BSC-4FA2**

Parameters:	Rated Voltage:	230V~
	Rated Frequency:	50Hz
	Rated Power:	1050W for BSC-3FA2 1150W for BSC-4FA2
	Protection Class:	Class I
	Degree of Protection:	IP X0
	Class of MSC:	II

Tested according to: EN 61010-1:2010/A1:2019

This Attestation of Conformity is issued on a voluntary basis according to the Low Voltage Directive 2014/35/EU relating to electrical equipment designed for use within certain voltage limits. It confirms that the listed equipment complies with the principal protection requirements of the directive and is based on the technical specifications applicable at the time of issuance. It refers only to the particular sample submitted for conformity assessment. For details see: www.tuvsud.com/ps-cert

Test report no.: 64111200354102**Date,** 2025-08-04
(Bin Li)

Page 1 of 1

This Attestation does not replace the regulatory EU Declaration of Conformity (DoC) and does not allow for CE marking. After preparation of the necessary documentation and establishing compliance to requirements of all applicable directives, the manufacturer may sign a DoC and apply the CE marking. The DoC is issued under the sole responsibility of the manufacturer.