ZERTIFIKAT ◆ CERTIFICATE ◆ 認證證書 ◆ СЕРТИФИКАТ ◆ CERTIFICADO ◆ CERTIFICAT

Product Service

C E R T I F I C A T E

No. Z1 083766 0006 Rev. 01

Holder of Certificate: **Jinan Biobase Biotech Co., Ltd.**
North Side of Jiwang Road
Mingshui Economic Development District
250200 Zhangqiu, Jinan City
PEOPLE'S REPUBLIC OF CHINA

Certification Mark:

Product: **Laboratory Equipment (Microbiological Safety Cabinet)**

Model(s): **BSC-3FA2, BSC-4FA2**

Parameters:

Rated Voltage:	230V~
Rated Frequency:	50Hz
Rated Power:	1050W for BSC-3FA2 1150W for BSC-4FA2
Protection Class:	Class I
Degree of Protection:	IP X0
Class of MSC:	II

Brand Name: **BIOBASE**

Tested according to: EN 61010-1:2010/A1:2019
EN 12469:2000

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition, the certification holder must not transfer the certificate to third parties. This certificate is valid until the listed date, unless it is cancelled earlier.
All applicable requirements of the Testing, Certification, Validation and Verification Regulations of TÜV SÜD Group have to be complied. For details see: www.tuvsud.com/ps-cert

Test report no.: 64111200354102
Valid until: 2030-08-03

Date, 2025-08-04

(Bin Li)

TÜV SÜD Product Service GmbH • Certification Body • Ridlerstraße 65 • 80339 Munich • Germany